

# Schedule of Accreditation



Organisation Name	Endoval Ltd.
Trading As	Endoval Limited
INAB Reg No	346T
Contact Name	Daniel Parra
Address	Unit B1 Hume Centre, Parkwest Industrial Park, Nangor Road, Dublin, D12 P3V9
Contact Phone No	01 6205429
Email	d.parra@endoval.ie
Website	<a href="http://www.endoval.ie">http://www.endoval.ie</a>
Accreditation Standard	EN ISO/IEC 17025 T
Standard Version	2017
Date of award of accreditation	07/12/2016
Scope Classification	Biological and veterinary testing
Services available to the public <sup>1</sup>	Yes

<sup>1</sup> Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered		
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)		
	Name	Address
1	Head Office	Unit B1 Hume Centre, Parkwest Industrial Park, Nangor Road, Dublin, D12 P3V9

# Scope of Accreditation

## Head Office

### Biological and Veterinary Testing

Category: A

Biology/veterinary field - Tests	Test name	Technique	Matrix	Equipment	Std. reference
803 Culture of organisms in liquid or agar based culture media with visual or instrument monitoring for growth - .01 Culture of bacteria	Pseudomonas aeruginosa (Presence/Absence)	IDEXX bacterial enzyme detection technology.	Low-ionic strength waters including washer-disinfectant final rinse waters and reverse osmosis waters.	IDEXX Pseudalert	I.S. EN ISO 15883-4:2018. section 6.3 HTM 01-06 Part E: Testing Methods. sections 6.91 to 6.96 Method SOP 030
	Total Viable Count onto TSA	Membrane Filtration	Low-ionic strength waters including washer-disinfectant final rinse waters and reverse osmosis waters.	Membrane Filtration	I.S. EN ISO 15883-1:2009. section 6.4 I.S. EN ISO 15883-4:2018. Annex E HTM 01-06 Part E Testing Methods. section 6.63 to 6.83 Method SOP 002
803 Culture of organisms in liquid or agar based culture media with visual or instrument monitoring for	Mycobacteria onto Middlebrook 7H10 media		Low-ionic strength waters including washer-disinfectant final rinse waters and reverse	Membrane Filtration	I.S. EN ISO 15883-4:2018. Annex E

growth - .03 Culture of mycobacteria			osmosis waters.		HTM 01-06 Part E: Testing Methods. sections 6.84 to 6.90 Method SOP 003
820 Miscellaneous	Detection and enumeration of microorganisms (bacteria, fungi and yeast) on TSA 55mm agar plates with lecithin & tween neutralisers	Culture of microorganisms on agar plates with visual monitoring of growth	Surfaces	Incubator at 30C	IS EN 16442:2015  QPSD-D-080-1  SOP 015
	Detection and enumeration of microorganisms (bacteria, fungi and yeast) on TSA 90mm agar plates	Culture of microorganisms on agar plates with visual monitoring of growth  SOP 015	Air	Incubator at 30C	IS EN 16442:2015  QPSD-D-080-1
	Detection and enumeration of microorganisms (bacteria, fungi and yeasts) in endoscope lumens	Recovery and culture of microorganisms on R2A agar plates with visual monitoring of growth	Autoclaved endoscope cleaning brushes	Class II microbiological cabinet Incubator at 30C Filtration station Ultrasonic bath	QPSD-D-080-1 page 68 Method SOP 013
822 Detection of Bacterial Endotoxins - .01 Kinetic Turbidimetric, CE marked commercial systems	Quantification of bacterial endotoxins in waters	Bacterial endotoxin determination using turbidimetric LAL assay	Low ionic strength waters including washer-disinfector final rinse waters, reverse osmosis waters, steam condensate, boiler feed water and dialysis water	Pyros Kinetix Flex Reader	IS EN ISO 15883-1:2009 HTM 01-01 Part C Steam Sterilization IS EN 285:2015 ISO 13959:2014 Method SOP 060